

Supplementary Information: Online Resource 4

Bland-Altman Bias 95% LOA			
	<i>n</i> =	Amplitude (Skin:GFE)	Peak time (Skin - GFE, ms)
<i>DA 0.01 b-wave</i>	39	0.19 <i>0.11 – 0.27</i>	-3.8 <i>-13 – 5.1</i>
<i>DA 3 a-wave</i>	41	0.21 <i>0.12 – 0.29</i>	-0.3 <i>-1.9 – 1.0</i>
<i>DA 3 b-wave</i>	41	0.19 <i>0.11 – 0.27</i>	-3.2 <i>-11 – 4.6</i>
<i>DA 10 a-wave</i>	41	0.20 <i>0.12 – 0.29</i>	-0.9 <i>-2.5 – 0.6</i>
<i>DA 10 b-wave</i>	41	0.19 <i>0.10 – 0.28</i>	0.9 <i>0.8 – 1.1</i>
<i>LA 30 Hz peak</i>	41	0.24 <i>0.14 – 0.34</i>	-0.3 <i>-1.0 – 0.4</i>
<i>LA 3 a-wave</i>	41	0.20 <i>0.11 – 0.29</i>	-0.8 <i>-3.1 – 1.5</i>
<i>LA 3 b-wave</i>	41	0.23 <i>0.13 – 0.32</i>	-0.5 <i>-1.3 – 0.3</i>

Table shows Bland-Altman analysis of skin electrode ERGs recorded simultaneously with gold foil ERGs from the right eye

“ISCEV standard full-field ERG reference limits from 407 healthy subjects, derived from transference and validation of reference data between electrode types and centres.” *Documenta Ophthalmologica*. RA Baker¹, SM Leo^{1,2}, WIN Clowes¹, I Chow³, X Jiang^{2,3}, AL Georgiou^{1,2}, A Calcagni¹, CJ Hammond³, MM Neveu^{1,2}, OA Mahroo^{1,2,3}, AG Robson^{1,2}. Affiliations: 1. Moorfields Eye Hospital NHS Foundation Trust. 2. UCL Institute of Ophthalmology, London. 3. St Thomas’ Hospital, London. Corresponding author e-mail: anthony.robson3@nhs.net