

Pipetting conditions	Average value	Standard deviation	Coefficient of variation	Number of samples
Manual transfer	5.56	0.11	0.019	12
Robotic transfer	5.29	0.15	0.028	96