

A.						
Lam	QT 0XSG		8X Gain			
(fg)	r2	E _{max}	ΔE	F _{max}	F _o	OCF
1,000	0.9975	98.5%	9.51E-05	10,365	3.04E-03	983
1,000	0.9976	96.8%	9.75E-05	9,922	3.39E-03	1095
100	0.9993	96.6%	1.02E-04	9,493	3.33E-04	1078
1,000	0.9983	92.5%	1.12E-04	8,289	4.59E-03	1483
100	0.9961	99.6%	1.21E-04	8,244	2.14E-04	693
1,000	0.9995	94.4%	1.22E-04	7,768	3.77E-03	1218
Av.:	0.9981	96.4%	1.08E-04	9,014		1092
SD/CV:	0.0013	2.6%	10.8%	11.7%		23.9%

B.						
Lam	QT 0.2XSG		8X Gain			
(fg)	r2	E _{max}	ΔE	F _{max}	F _o	OCF
100	0.9991	99.1%	3.89E-05	25,464	3.68E-04	1188
100	0.9993	99.2%	3.45E-05	28,727	4.14E-04	1338
100	0.9991	99.1%	3.89E-05	25,464	3.68E-04	1188
100	0.9984	97.7%	4.08E-05	23,927	3.78E-04	1222
100	0.9995	100.5%	4.60E-05	21,861	2.58E-04	836
Av.:	0.9991	99.1%	3.98E-05	25,089		1155
SD/CV:	0.0004	1.0%	10.4%	10.0%		16.3%

C.						
Lam	DyNa 0.5XSG		2X Gain			
(fg)	r2	E _{max}	ΔE	F _{max}	F _o	OCF
100	0.9976	93.4%	8.86E-05	10,540	3.35E-04	1085
100	0.9982	96.9%	8.93E-05	10,845	1.71E-04	553
100	0.9995	97.0%	8.67E-05	11,187	1.93E-04	625
100	0.9985	94.5%	9.88E-05	9,564	2.35E-04	761
100	0.9996	95.0%	9.42E-05	10,088	2.38E-04	771
100	0.9972	94.8%	9.19E-05	10,317	2.00E-04	645
100	0.9984	96.8%	9.51E-05	10,179	1.78E-04	574
100	0.9992	95.7%	1.09E-04	8,798	1.94E-04	626
100	0.9992	98.3%	8.95E-05	10,976	1.56E-04	504
100	0.9988	93.0%	1.47E-04	6,309	2.76E-04	891
Av.:	0.9987	95.5%	9.90E-05	9,880		704
SD/CV:	0.0008	1.7%	18.4%	14.6%		25.2%

D.						
Lam	DyNa 1.5XSG		2X Gain			
(fg)	r2	E _{max}	ΔE	F _{max}	F _o	OCF
100	0.9989	92.1%	3.26E-05	28,278	2.49E-04	806
100	0.9998	92.0%	3.19E-05	28,839	2.84E-04	918
100	0.9995	94.0%	3.92E-05	23,949	2.37E-04	765
100	0.9979	93.7%	4.05E-05	23,147	2.01E-04	650
100	0.9974	93.6%	3.72E-05	25,185	2.33E-04	753
Av.:	0.9987	93.1%	3.63E-05	25,880		778
SD/CV:	0.0010	0.9%	10.7%	9.9%		12.4%

E.						
Lam	FV 2.0XSG		1X Gain			
(fg)	r2	E _{max}	ΔE	F _{max}	F _o	OCF
1,000	0.9970	96.3%	8.06E-05	11,954	1.12E-03	362
1,000	0.9999	94.9%	8.74E-05	10,861	1.51E-03	487
1,000	0.9993	95.2%	6.56E-05	14,524	1.14E-03	368
1,000	0.9994	96.4%	7.85E-05	12,289	6.86E-04	222
1,000	0.9978	95.0%	1.00E-04	9,499	7.42E-04	240
1,000	0.9990	94.2%	7.83E-05	12,031	1.20E-03	388
Av.:	0.9988	95.4%	8.17E-05	11,860		345
SD/CV:	0.0011	0.9%	14.0%	14.1%		28.8%
Overall CV Average:						21.3%