

Supporting Information

Contents:

1. CCK8 assay
2. Compound **5d** induces apoptosis in HepG2 cells
3. Molecular Docking Study
4. ^1H NMR spectra, ^{13}C NMR, IR and HRMS spectra of compounds **5a-5l**,
6a-6j and **7a, 7b**


1. CCK8 assay the compounds **5a-5l**, **6a-6j** and **7** were test their antitumor activity using the CCK8 assay as described previously. HepG2 cells were cultivated in a density in 96-well plates (5000 cells/well). After incubation for 24 h, cells were treated with compounds for 48 h at a final compound concentration of 0.1-50 $\mu\text{mol/L}$. Observing the absorbance at 490 nm with a microplate reader, the cell inhibition rate = $((1-A_{\text{experiment}})/A_{\text{control}}) \times 100\%$, and calculating the IC_{50} subsequently.

2. Compound **5d** induces apoptosis in HepG2 cells the apoptotic effect of compound **5d** was carried out with propidium iodide (PI) and FITC-annexin-V stain. According to previously reported method and guidance of instructions, HepG2 cells were treated with compound **5d** at 0, 5.0, 10.0, 15.0 $\mu\text{mol/L}$, respectively, for 48 h, and then analyzed by flow cytometry assay.


3. Molecular Docking Study

The three-dimension (3D) structures of all compounds were prepared by SYBYL 7.0 followed by 2,000-steps steepest descent minimization and 2,000-steps conjugate gradient minimization, respectively. The crystal structure of EGFR in complex with Erlotinib (FDA approved drug) was obtained from Protein Data Bank (PDB ID: 1M17). The Erlotinib was deleted from the complex structure and the location of it was used to define the active site. The hydrogen of the receptor were added by using Discovery Studio 4.0. The GOLD 3.0 was used to dock compound **5d** into the active center. The radius of active site was set to 10 Å by using a genetic algorithm (GA) with 300 runs. The top rank conformation of the two compounds were selected as the representative. The docking score is calculated using Autodock Score and expressed in terms of binding free energies (ΔG_{cal} , Kcal/mol). The experimental binding free energies (ΔG_{exp} , Kcal/mol) using the equation $\Delta G = RT \ln IC_{50}$.

4. ^1H NMR spectra, ^{13}C NMR, IR and HRMS spectra of compounds **5a-5l**, **6a-6j** and **7a, 7b**



^1H NMR **5a**



^{13}C NMR 5a

Elemental Composition Report


Page 1

Single Mass Analysis

Tolerance = 5.0 mDa / DBE: min = -1.5, max = 50.0

Element prediction: Off

Number of isotope peaks used for i-FIT = 3



Monoisotopic Mass, Even Electron Ions

972 formula(e) evaluated with 1 results within limits (up to 50 best isotopic matches for each mass)


Elements Used:

C: 26-26 H: 28-28 N: 0-100 O: 0-100 Na: 0-1

7

240331-2-8-8 11 (0.076)


1: TOF MS ES+
2.01e+005




Minimum: -1.5
Maximum: 5.0 10.0 50.0

Mass	Calc. Mass	mDa	PPM	DBE	i-FIT	Norm	Conf (%)	Formula
490.2095	490.2090	0.5	1.0	15.5	163.9	n/a	n/a	C ₂₆ H ₂₈ N ₅ O ₅


HRMS m/z 5a




4. ^1H NMR spectra, ^{13}C NMR and HRMS spectra of compounds **5a-5l**, **6a-6j** and **7a-7b**.



^1H NMR **5b**





Elemental Composition Report

Single Mass Analysis

Tolerance = 5.0 mDa / DBE: min = -1.5, max = 50.0

Element prediction: Off

Number of isotope peaks used for i-FIT = 3

Monoisotopic Mass, Even Electron Ions

972 formula(e) evaluated with 1 results within limits (up to 50 best isotopic matches for each mass)



Elements Used:

C: 26-26 H: 28-28 N: 0-100 O: 0-100 Na: 0-1

7

240331-2-8-7 13 (0.083)

Page 1






1: TOF MS ES+
4.96e+005

Minimum: -1.5
Maximum: 5.0 10.0 50.0


Mass	Calc. Mass	mDa	PPM	DBE	i-FIT	Norm	Conf (%)	Formula
490.2093	490.2090	0.3	0.6	15.5	193.7	n/a	n/a	C26 H28 N5 O5

HRMS m/z 5b





¹H NMR 5c



^{13}C NMR 5c


Elemental Composition Report

Single Mass Analysis

Tolerance = 5.0 mDa / DBE: min = -1.5, max = 50.0

Element prediction: Off

Number of isotope peaks used for i-FIT = 3



Page 1

Monoisotopic Mass, Even Electron Ions

1030 formula(e) evaluated with 1 results within limits (up to 50 best isotopic matches for each mass)


Elements Used:

C: 26-26 H: 28-28 N: 0-100 O: 0-100 Na: 0-1

7

240331-2-8-9 10 (0.072)


1: TOF MS ES+
5.80e+005




Minimum: -1.5
Maximum: 5.0 10.0 50.0


Mass	Calc. Mass	mDa	PPM	DBE	i-FIT	Norm	Conf (%)	Formula
506.2043	506.2040	0.3	0.6	15.5	210.6	n/a	n/a	C26 H28 N5 O6

HRMS m/z 5c





^1H NMR 5d



^{13}C NMR 5d

Elemental Composition Report

Single Mass Analysis

Tolerance = 5.0 mDa / DBE: min = -1.5, max = 50.0

Element prediction: Off

Number of isotope peaks used for i-FIT = 3

Monoisotopic Mass, Even Electron Ions


3587 formula(e) evaluated with 1 results within limits (up to 50 best isotopic matches for each mass)

Elements Used:

C: 25-25 H: 5-25 N: 0-100 O: 0-100 Na: 0-2 Cl: 1-3

2


240503-3-16 19 (0.196)


1: TOF MS ES+
4.79e+006


Minimum: -1.5
Maximum: 5.0 10.0 50.0


Mass	Calc. Mass	mDa	PPM	DBE	i-FIT	Norm	Conf (%)	Formula
510.1547	510.1544	0.3	0.6	15.5	651.7	n/a	n/a	C ₂₅ H ₂₅ N ₅ O ₅ Cl ₁

HRMS m/z 5d









^{13}C NMR 5e

Elemental Composition Report

Single Mass Analysis

Tolerance = 5.0 mDa / DBE: min = -1.5, max = 50.0

Element prediction: Off

Number of isotope peaks used for i-FIT = 3

Monoisotopic Mass, Even Electron Ions

3996 formula(e) evaluated with 1 results within limits (up to 50 best isotopic matches for each mass)


Elements Used:

C: 25-25 H: 24-24 N: 0-100 O: 0-100 Na: 0-2 F: 1-3

2

240503-3-18 19 (0.196)


1: TOF MS ES+
9.06e+005





Minimum: -1.5
Maximum: 50.0

Mass	Calc. Mass	mDa	PPM	DBE	i-FIT	Norm	Conf (%)	Formula
516.1659	516.1659	0.0	0.0	15.5	413.2	n/a	n/a	C ₂₅ H ₂₄ N ₅ O ₅ Na F


HRMS m/z 5e









^1H NMR 5f




Sample Name		Position	Vial 51	Instrument Name	Instrument 1	User Name	
Inj Vol	2	InjPosition		SampleType	Sample	IRM Calibration Status	Success
Data Filename	5k.d	ACQ Method	CYF20201112-NORMAL.m	Comment		Acquired Time	3/17/2025 11:19:23




HRMS m/z 5f





¹H NMR 5g



^{13}C NMR 5g

Elemental Composition Report

Single Mass Analysis

Tolerance = 5.0 mDa / DBE: min = -1.5, max = 50.0

Element prediction: Off

Number of isotope peaks used for i-FIT = 3

Monoisotopic Mass, Even Electron Ions

1902 formula(e) evaluated with 1 results within limits (up to 50 best isotopic matches for each mass)


Elements Used:

C: 27-27 H: 30-30 N: 0-100 O: 0-100 Na: 0-3

14

240612-1-z-28 21 (0.147)


1: TOF MS ES+
8.79e+005




Minimum: -1.5
Maximum: 5.0 10.0 50.0


Mass	Calc. Mass	mDa	PPM	DBE	i-FIT	Norm	Conf (%)	Formula
504.2252	504.2247	0.5	1.0	15.5	681.2	n/a	n/a	C ₂₇ H ₃₀ N ₅ O ₅

HRMS m/z 5g






^1H NMR 5h



¹³C NMR 5h



Single Mass Analysis

Tolerance = 5.0 mDa / DBE: min = -1.5, max = 50.0

Element prediction: Off

Number of isotope peaks used for i-FIT = 3

Monoisotopic Mass, Even Electron Ions

3587 formula(e) evaluated with 1 results within limits (up to 50 best isotopic matches for each mass)


Elements Used:

C: 25-25 H: 5-25 N: 0-100 O: 0-100 Na: 0-2 Cl: 1-3

2

240503-3-15 22 (0.222)


1: TOF MS ES+
4.76e+006




Minimum: -1.5
Maximum: 5.0 10.0 50.0

Mass	Calc. Mass	mDa	PPM	DBE	i-FIT	Norm	Conf (%)	Formula
510.1546	510.1544	0.2	0.4	15.5	694.8	n/a	n/a	C25 H25 N5 O5 Cl

HRMS m/z 5h



^1H NMR 5i



Elemental Composition Report

Single Mass Analysis

Tolerance = 5.0 mDa / DBE: min = -1.5, max = 50.0

Element prediction: Off

Number of isotope peaks used for i-FIT = 3

Monoisotopic Mass, Even Electron Ions

1790 formula(e) evaluated with 1 results within limits (up to 50 best isotopic matches for each mass)


Elements Used:

C: 26-26 H: 28-28 N: 0-100 O: 0-100 Na: 0-3

14

240612-1-z-32 29 (0.189)


1: TOF MS ES+
2.52e+005





Minimum: -1.5
Maximum: 5.0 10.0 50.0

Mass	Calc. Mass	mDa	PPM	DBE	i-FIT	Norm	Conf (%)	Formula
490.2095	490.2090	0.5	1.0	15.5	538.2	n/a	n/a	C26 H28 N5 O5


HRMS m/z 5i







¹H NMR 5j



^{13}C NMR 5j

Elemental Composition Report

Single Mass Analysis

Tolerance = 5.0 mDa / DBE: min = -1.5, max = 50.0

Element prediction: Off

Number of isotope peaks used for i-FIT = 3

Monoisotopic Mass, Even Electron Ions

913 formula(e) evaluated with 1 results within limits (up to 50 closest results for each mass)


Elements Used:

C: 25-25 H: 25-25 N: 0-100 O: 0-100 Na: 0-1 F: 1-1

9

240629-4-17 10 (0.089)


1: TOF MS ES+
3.23e+005





Minimum: -1.5
Maximum: 5.0 10.0 50.0

Mass	Calc. Mass	mDa	PPM	DBE	i-FIT	Norm	Conf (%)	Formula
494.1844	494.1840	0.4	0.8	15.5	468.8	n/a	n/a	C ₂₅ H ₂₅ N ₅ O ₅ F


HRMS m/z 5j









^1H NMR 5k







User Spectra




HRMS m/z 5k





¹H NMR 5I



¹³C NMR 51

Elemental Composition Report

Single Mass Analysis

Tolerance = 5.0 mDa / DBE: min = -1.5, max = 50.0

Element prediction: Off

Number of isotope peaks used for i-FIT = 3

Monoisotopic Mass, Even Electron Ions

972 formula(e) evaluated with 1 results within limits (up to 50 closest results for each mass)


Elements Used:

C: 26-26 H: 28-28 N: 0-100 O: 0-100 Na: 0-1

9

240629-4-18 12 (0.100)


1: TOF MS ES+
1.67e+004





Minimum: -1.5
Maximum: 5.0 10.0 50.0

Mass	Calc. Mass	mDa	PPM	DBE	i-FIT	Norm	Conf (%)	Formula
490.2093	490.2090	0.3	0.6	15.5	305.6	n/a	n/a	C26 H28 N5 O5


HRMS m/z 51








¹H NMR 6a





^{13}C NMR 6a

Sample Name		Position	Vial 51	Instrument Name	Instrument 1	User Name	
Inj Vol	2	InjPosition		SampleType	Sample	IRM Calibration Status	Success
Data Filename	6b.d	ACQ Method	CYF20201112-NORMAL.m	Comment		Acquired Time	3/14/2025 4:59:54 P




HRMS m/z **6a**






^1H NMR 6b





^{13}C NMR 6b

Sample Name		Position	Vial 51	Instrument Name	Instrument 1	User Name	
Inj Vol	2	InjPosition		SampleType	Sample	IRM Calibration Status	Success
Data Filename	6c.d	ACQ Method	CYF20201112-NORMAL.m	Comment		Acquired Time	3/17/2025 3:01:18 P




HRMS m/z **6b**





^1H NMR 6c



^{13}C NMR **6c**

Elemental Composition Report

Page 1

Single Mass Analysis

Tolerance = 5.0 mDa / DBE: min = -1.5, max = 50.0

Element prediction: Off

Number of isotope peaks used for i-FIT = 3

Monoisotopic Mass, Even Electron Ions

5652 formula(e) evaluated with 1 results within limits (up to 50 closest results for each mass)


Elements Used:

C: 26-26 H: 25-25 N: 0-100 O: 0-100 Na: 0-4 K: 0-2

1

240905-6-HYG-Z-54 31 (0.219)


1: TOF MS ES+
6.22e+004




Minimum: -1.5
Maximum: 5.0 10.0 50.0


Mass	Calc. Mass	mDa	PPM	DBE	i-FIT	Norm	Conf (%)	Formula
494.1805	494.1804	0.1	0.2	16.5	666.4	n/a	n/a	C26 H25 N5 O4 Na

HRMS m/z 6c





^1H NMR **6d**



^{13}C NMR 6d

Elemental Composition Report

Page 1

Single Mass Analysis

Tolerance = 5.0 mDa / DBE: min = -1.5, max = 50.0

Element prediction: Off

Number of isotope peaks used for i-FIT = 3

Monoisotopic Mass, Even Electron Ions


5158 formula(e) evaluated with 1 results within limits (up to 50 closest results for each mass)

Elements Used:


C: 26-26 H: 26-26 N: 0-100 O: 0-100 Na: 0-4 K: 0-2

1

240905-6-HYG-Z-55 11 (0.094)




1: TOF MS ES+
1.77e+006




Minimum: -1.5
Maximum: 5.0 10.0 50.0


Mass	Calc. Mass	mDa	PPM	DBE	i-FIT	Norm	Conf (%)	Formula
472.1989	472.1985	0.4	0.8	16.5	988.9	n/a	n/a	C ₂₆ H ₂₆ N ₅ O ₄

HRMS m/z 6d





^1H NMR 6e



^{13}C NMR 6e

Elemental Composition Report

Single Mass Analysis

Tolerance = 5.0 mDa / DBE: min = -1.5, max = 50.0

Element prediction: Off

Number of isotope peaks used for i-FIT = 3

Monoisotopic Mass, Even Electron Ions

5158 formula(e) evaluated with 1 results within limits (up to 50 closest results for each mass)


Elements Used:

C: 26-26 H: 26-26 N: 0-100 O: 0-100 Na: 0-4 K: 0-2


1

240905-6-HYG-Z-57 12 (0.100)

Page 1




1: TOF MS ES+
2.07e+006




Minimum: -1.5
Maximum: 5.0 10.0 50.0


Mass	Calc. Mass	mDa	PPM	DBE	i-FIT	Norm	Conf (%)	Formula
472.1988	472.1985	0.3	0.6	16.5	993.4	n/a	n/a	C26 H26 N5 O4

HRMS m/z 6e





^1H NMR **6f**



^{13}C NMR 6f

Elemental Composition Report

Page 1

Single Mass Analysis

Tolerance = 5.0 mDa / DBE: min = -1.5, max = 50.0

Element prediction: Off

Number of isotope peaks used for i-FIT = 3

Monoisotopic Mass, Even Electron Ions


9777 formula(e) evaluated with 1 results within limits (up to 50 closest results for each mass)

Elements Used:


C: 25-25 H: 22-22 N: 0-100 O: 0-100 Na: 0-4 Cl: 1-2 K: 0-2

1

240905-6-HYG-Z-58 23 (0.157)




1: TOF MS ES+
2.37e+006




Minimum: -1.5
Maximum: 5.0 10.0 50.0


Mass	Calc. Mass	mDa	PPM	DBE	i-FIT	Norm	Conf (%)	Formula
514.1262	514.1258	0.4	0.8	16.5	978.5	n/a	n/a	C ₂₅ H ₂₂ N ₅ O ₄ Na Cl

HRMS m/z 6f






^1H NMR **6g**





^{13}C NMR 6g

Sample Name		Position	Vial 51	Instrument Name	Instrument 1	User Name	
Inj Vol	2	InjPosition		SampleType	Sample	IRM Calibration Status	Success
Data Filename	71.d	ACQ Method	CYF20201112-NORMAL.m	Comment		Acquired Time	3/14/2025 9:06:22 P




HRMS m/z 6g






^1H NMR **6h**





^{13}C NMR 6h

Sample Name		Position	Vial 52	Instrument Name	Instrument 1	User Name	
Inj Vol	2	InjPosition		SampleType	Sample	IRM Calibration Status	Success
Data Filename	75.d	ACQ Method	CYF20201112-NORMAL.m	Comment		Acquired Time	3/17/2025 3:10:56 P




HRMS m/z 6h







^1H NMR 6i




¹³C NMR 6i




User Spectra




HRMS m/z **6i**





¹H NMR 6j



^{13}C NMR 6j

Elemental Composition Report

Page 1

Single Mass Analysis

Tolerance = 5.0 mDa / DBE: min = -1.5, max = 50.0

Element prediction: Off

Number of isotope peaks used for i-FIT = 3

Monoisotopic Mass, Even Electron Ions

5158 formula(e) evaluated with 1 results within limits (up to 50 closest results for each mass)


Elements Used:

C: 26-26 H: 26-26 N: 0-100 O: 0-100 Na: 0-4 K: 0-2

1

240905-6-HYG-Z-55 11 (0.094)


1: TOF MS ES+
1.77e+006




Minimum: -1.5
Maximum: 5.0 10.0 50.0

Mass	Calc. Mass	mDa	PPM	DBE	i-FIT	Norm	Conf (%)	Formula
472.1989	472.1985	0.4	0.8	16.5	988.9	n/a	n/a	C26 H26 N5 O4

HRMS m/z 6j



¹H NMR 7a



^{13}C NMR 7a

IRM Calibration Status
Comment

Success


DA Method

TQQ-18-23-20250109.m


Sample Group
Acquisition SW
Version

6200 series TOF/6500 series
Q-TOF B.05.00 (B5042.2)


Info.




User Spectra




HRMS m/z 7a






^1H NMR **7b**



^{13}C NMR 7b

Sample Name		Position	Vial 52	Instrument Name	Instrument 1	User Name	
Inj Vol	2	InjPosition		SampleType	Sample	IRM Calibration Status	Success
Data Filename	ZYG--250318-2.d	ACQ Method	CYF20201112-NORMAL.m	Comment		Acquired Time	3/18/2025 12:04:35



HRMS m/z **7b**

