

Self-sampling feasibility results according to randomization arm

Variable	Upper Arm (n = 25)	Finger Prick (n = 25)
Patients succeeding at first attempt, n (%)	20 (80.0)	21(84.0)
Patients succeeding at second attempt, n (%)	4 (16.0)	4 (16.0)
Patients not succeeding at all, n (%)	1 (4.0)	0 (0.0)
Patients requiring assistance, n (%)	8 (32.0)	7 (28.0)
Reasons for assistance ^a , n		
Opening kit	3	x
Pressing button	5	x
Removing button cap (UA) / twisting off the tab (FP)	1	6
Removing collection tube	3	x
Collecting blood	x	3
Volume of blood collected (µl), mean ± SD	106.2 ± 60.6	118.8 ± 74.7
Patients with sufficient Blood for analysis of all three biomarkers, n (%)	10 (40.0)	12 (48.0)

^aMultiple reasons could be recorded per patient