

Additional file 5




Figure S4.

Spiked recovery. The results of the spiked recovery test. Purified mono-Ln- γ 2 protein was added to 3 normal urine specimens to achieve final concentrations of 50 and 1,000 pg/mL. The samples were tested using the CLIA and the recovery rate was calculated for the purified mono-Ln- γ 2. The recovery rates from three healthy urine specimens were 82.3–90.3% for a final concentration of 50 pg/mL and 86.1–96.2% for a final concentration of 1,000 pg/mL.