

Table EV1: Specificity and sensitivity of the RT-LAMP assay. Sensitivity (for each Ct bin) and specificity of the RT-LAMP test with respect to the RT-qPCR results, as derived from the results shown in Fig. 3.

Ct range	Sensitivity [%]	95% CI [%]	Specificity [%]	95% CI [%]
0-25	100.0	95.9 - 100.0	99.0%	97.6 - 99.6
>25-30	100.0	95.4 - 100.0		
>30-34	100.0	93.7 - 100.0		
>34-36	83.3	61.8 - 94.5		